



# Certificate of Conformity

## Grey Scale for Assessing Change in Colour (long Version)

**Standard** DIN EN 20105-A02 / ISO 105-A02  
**Batch** A02G0818  
**Test Report** WIWeB R1/0000023777-15 dated 30.10.2018  
**Result** *(Extract from the test report)*

Grade	Y-Value ISO 105-A02	$\Delta E$ ISO 105-A02	Tolerance	Result
5	12 +/- 1	0	0,2	pass
4-5	12 +/- 1	0,8	$\pm 0,2$	pass
4	12 +/- 1	1,7	$\pm 0,3$	pass
3-4	12 +/- 1	2,5	$\pm 0,35$	pass
3	12 +/- 1	3,4	$\pm 0,4$	pass
2-3	12 +/- 1	4,8	$\pm 0,5$	pass
2	12 +/- 1	6,8	$\pm 0,6$	pass
1-2	12 +/- 1	9,6	$\pm 0,7$	pass
1	12 +/- 1	13,6	$\pm 1,0$	pass

DEK (Deutsche Echtheitskommission - German Colour Fastness Committee) is the manufacturer of the above-mentioned Gray Scale for Assessing Change in Colour and regularly checks production and stock for conformity to standards.

This certificate confirms the conformity to standards of the product obtained from DEK with the above batch number with regard to the requirements of DIN EN 20105-A02.

This certificate is only valid with a corresponding proof of delivery from DEK or an official DEK sales partner.

Alexander Dietel  
President of DEK

Troisdorf, 4<sup>th</sup> November 2021